

# HVP-40 / 50 DC High Voltage Probe To DMM

## SPECIFICATIONS:

Model	HVP - 40	HVP - 50
Ratio	1000 : 1 (1 KV = DMM 1 V Reading Value )	1000 : 1 (1 KV = DMM 1 V Reading Value )
Input Impedance	1000M $\Omega$ $\pm$ 7%	500M $\Omega$ $\pm$ 2%
Temperature Coefficient	$\leq$ 200 ppm / $^{\circ}$ C	$\leq$ 100 ppm / $^{\circ}$ C
Max. Input Voltage	DC: 0 ~ 40 KV AC: 0 ~ 28 KV (50 / 60 Hz)	DC: 0 ~ 50 KV AC: 0 ~ 35 KV (50 / 60 Hz)
Accuracy	DC: $\pm$ 1% to 20 KV, $\pm$ 2% to 40 KV AC: $\pm$ 5% to 28 KV (50 / 60 Hz only)	DC: $\pm$ 1% to 40 KV, $\pm$ 2% to 50 KV AC: $\pm$ 2% to 28 KV, $\pm$ 3% to 35 KV (50 / 60 Hz only)
Max. Loading Current	40 $\mu$ A at 40 KV DC	0.1 mA at 50 KV DC
Max. Loading Power	1.6 Watt	5 Watt
Ground Load Length	90 cm (3 ft)	90 cm (3 ft)
Output Cable Length	100 cm (3.3 ft)	100 cm (3.3 ft)
Dimensions / Weight	340 mm x 80 $\varnothing$ / About 230 g	340 mm x 80 $\varnothing$ / About 260 g
Operating Temperature	0 $^{\circ}$ C ~ + 50 $^{\circ}$ C	0 $^{\circ}$ C ~ + 50 $^{\circ}$ C
Operating Temperature	-20 $^{\circ}$ C ~ + 70 $^{\circ}$ C	-20 $^{\circ}$ C ~ + 70 $^{\circ}$ C

### ⚠ WARNING ⚠

- PLEASE CONNECT TO GROUND LEAD AND CABLES PROBABLY AND DOING THE MEASUREMENTS.
- PLEASE PUT PROBE ON THE ISOLATED TABLE WHEN DOING THE MEASUREMENTS.
- PLEASE DO NOT TOUCH PROBE AND CABLES WHEN DOING THE MEASUREMENTS.

### ACCESSORIES:

(Test CRT for H.V.)



# **PINTEK**

## **Pintek Electronics Co., Ltd**

2F No.32 Tai-Pen Rd., Shulin Dist., New Taipei City, Taiwan

Tel: 886-2-86859562~3 ; 886-2-86861898

Fax: 886-2-86852081 ; 886-2-86861114

<http://www.pintek.com.tw>

E-Mail : [pintek@pintek.com.tw](mailto:pintek@pintek.com.tw) ; [pintek@ms26.hinet.net](mailto:pintek@ms26.hinet.net)

SALES AGENT