


LCR-900 High Precision LCR Meter

(100KHz With USB Interface)



SPECIFICATIONS:

TEST FREQUENCY		100Hz to 100KHz (5 Range)
BASIC ACCURACY		0.3%
TEST SIGNAL LEVEL		0.6Vrms
DISPLAY RANGE	Impedance R	0.000 Ω to 9999 MΩ
	Capacitance C	0.000 pF to 9999 F
	Inductance L	0.000 uH to 9999 KH
	Equivalent Series Resistance ESR	0.000 Ω to 9999 Ω
	Equivalent Parallel Resistance Rp	0.000 Ω to 9999 Ω
	Quality Factor Q	0.000 to 9999
	Dissipation Factor D	0.000 to 9999
Phase Angle Θ		-90° to 90°
MEASUREMENT PARAMETER		Z, L, C, DCR, ESR, D, Q, Θ
OPERATION SELECT		AUTO/MANUAL
DISPLAY MODE		Value, Δ, Δ%, Sorting, Sorting Range
INTERFACE(OPTION)		USB
POWER SOURCE		115/230 ± 10%, 50/60Hz

ACCESSORIES:



PINTEK

Pintek Electronics Co., Ltd

2F No.32 Tai-Pen Rd., Shulin Dist., New Taipei City, Taiwan

Tel: 886-2-86859562~3 ; 886-2-86861898

Fax: 886-2-86852081 ; 886-2-86861114

http://www.pintek.com.tw

E-Mail : pintek@pintek.com.tw ; pintek@ms26.hinet.net

SALES AGENT